

Calibration Services for Analog Pressure Barometers and Transmitters

Please fill in the model, serial number and order code with the selected options:

Model:	Serial no:	Code: 223046			A		A	A	A
Calibration class	ISO/IEC 17025 Accredited calibration		A						
	ISO 10012 Compliant calibration		B						
Pressure calibration	510, 600, 700, 800, 900, 1000 and 1060 hPa		A						
NOTE 1	610, 700, 800, 900, 1000 and 1060 hPa		B						
	810, 850, 900, 950, 1000 and 1060 hPa		C						
	510, 600, 700, 800, 900, 1000 and 1100 hPa		D						
	610, 700, 800, 900, 1000 and 1100 hPa		E						
	810, 850, 900, 950, 1000, 1050 and 1100 hPa		F						
	910, 950, 1000, 1050 and 1100 hPa		G						
	510, 600, 700, 800, 900, 1000, 1100, 1200 and 1300 hPa		H						
	100, 200, 400, 600, 800, 1000, 1200 and 1300 hPa		I						
	810, 900, 1000, 1100, 1200 and 1300 hPa		J						
Calibration temperature	+23 °C (room temperature)		A						
Adjustment	No adjustment		A						
	Adjusted at 1000 hPa pressure (offset)		B						
	Adjusted for optimum performance		C						
Calibration method	Measurement from analog output		A						
Results reported, pressure	hPa		A						
Language of the Certificate	Certificate in English		A						

NOTE 1: Selected calibration range shall not exceed the measurement range of the transmitter.

This service is available to all analog pressure transmitters having output ranges between 0 to 10 VDC or 0 to 20 mADC in the pressure ranges listed in this order form. For transmitters made by other manufacturers than Vaisala, the operation and adjustment instructions of the transmitter shall be delivered with the transmitter.

Requested recalibration sticker (due-date):
YES/NO <input type="checkbox"/> Interval: <input type="text"/> months

Please fill in the end customer information for Certificate:

End customer	
Order received and approved	